

Nr. Crt.	Reactivi laborator	Specificatii de calitate	UM	Cantitate
	LOT 7			
1	Solutie standard pH 10 (25 °C), import	boric acid/potassium chloride/sodium hydroxide traceable to SRM from NIST and PTB pH 10.00 (25°C)	500 ml	2
2	Solutie standard pH 7 (25 °C), import	potassium dihydrogen phosphate/disodium hydrogen phosphate traceable to SRM from NIST and PTB pH 7.00 (25°C)	500 ml	2
3	Solutie standard pH 4 (25 °C), import	potassium hydrogen phthalate directly traceable to primary SRM from NIST/PTB pH (S) = 4.00 _s (25°C)	500 ml	2
4	Solutie standard azotat , 1000 mg/l, trasabila la SRM from NIST, MRC,import	Concentration _____ β NO ₃ ⁻ 990 - 1010 mg/l	500ml	1
5	Solutie standard cadmiu, 1000 mg/l, trasabila la SRM from NIST, pentru AAS, MRC,import	Concentration _____ β Cd 990 - 1010 mg/l	500ml	1
6	Solutie standard cianura, 1000 mg/l, trasabila la SRM from NIST, MRC,import	Concentration _____ β CN ⁻ 990 - 1010 mg/l	500ml	1
7	Solutie standard cupru 1000 mg/l, trasabila la SRM from NIST, MRC, import	Concentration _____ β Cu 990 - 1010 mg/l Determination method: Complexometric titration. (traceable to NIST - SRM 682) Accuracy of the method: +/- 2 mg/l	500 ml	1
8	Solutie standard TOC, hydrogen ftalat de potasiu 100 mg/l, import	(traceable to NIST - SRM 84k) Measurement uncertainty: +/-	100ml	1
9	Solutie TISAB II	Solutie pentru electrozii ISE de fluoruri	500 ml	2

10	Solutie standard KCl 0,01 mol/l, MR WTW 1,41 mS/cm	MRC Solutie de clorura de potasiu 0,01 mol/l , cu conductivitatea 1,413 mS/cm +- 0,020 mS/cm (25 °C)	500 ml	1
11	MRC Cianuri totale	Aprox 0,240 mg/l - 0.3 mg/l	fiole 2 ml	1
12	MRC Fenoli totali	0,06-5 mg/l	fiole 2 ml	1
13	MRC Anionic Surfactant-WP	0,6 mg/l	flacon 20 ml	1
14	MRC pH in matrice apa de suprafata sau apa uzata	5-10 unitati pH	250 ml	1
15	MRC duritate totala	Total hardness as CaCO3, domeniu 83-307 mg/l	250 ml	1
16	MRC azotati	0.1 mg/l NO3-N in H2O	flacon 100 ml	1
17	Test kit fluoruri, Spectroquant	domeniu : 0,1 - 2,5 mg/l, cat.1.17243	cutie cu 25 teste	1
18	Test kit detergenti, WTW	WTW stored Test Protocol 14697 for Ten-A (Surfactants Anionic)	cutie cu 25 fiole	1
19	Test kit cianura, Spectroquant	domeniu : 0,010- 0,500 mg/l CN	cutie	1
TOTAL				

Specificatiile din caietul de sarcini sunt minime si obligatorii. Vor fi luate in considerare numai ofertele care respecta aceste specificatii sau sunt superioare.

Intocmit :

R.U.C. ing. Alina Iacovliev

Alina